

T-CAL Series

Solid State Turbidity Calibration Kits

Product Description

Pyxis Lab has developed reusable solid-state turbidity calibration kits for rapid calibration of the Pyxis LT-73X Series inline ultra-low turbidity sensors. The LT-73X Series ultra-low turbidity sensors are factory calibrated on Formazin liquid standards. The T-CAL calibration kits represent an alternative to Formazin calibration for routine field calibration only.

To calibrate a LT-73X series inline sensor using Formazin, a large quantity of calibration solution must be added to the FR-100 flow reservoir. Perhaps, the most challenging aspect of calibrating an inline sensor is that zero-NTU water is extremely difficult to be prepared outside of a laboratory environment and may be very easily contaminated by particulate or air bubbles. The Pyxis solid-state calibration kits provide a consistent and reproducible method to calibrate Pyxis inline LT-73X series turbidity sensors, overcoming the challenges associated with the liquid standard calibration.



T-CAL Turbidity Calibration Kit

T-CAL Solid State Kit Use Method

The sensor should be removed from process flow, cleaned gently with a soft towel, inserted into the appropriate T-CAL adapter and wirelessly calibrated via the uPyxis Mobile or Desktop APP. Please refer to LT-73X Series



Turbidity Calibration Principals & Considerations

The standard turbidity values of the solid-state adapters are established in our laboratory by matching with their corresponding liquid Formazin standards prior to shipment. If the Pyxis sensor is purchased with the solid-state calibration kit, the turbidity values of the adapters are traceable to the Formazin standards. For applications that require accuracy better than 10% of the value and the solid-state calibration kit is not purchased with the sensor, the user can first calibrate the sensor with using Formazin standards and then measure and record the equivalent turbidity values of the solid-state adapters for future calibrations. For applications that do not require better than 10% accuracy, the user can simply calibrate the sensor to the original default values marked on the solid-state adapters.

The precision, resolution and the low detection limit are not affected by the calibration method, regardless of using the solid-state or liquid Formazin. The calibration only affects the turbidity sensor accuracy. The nature of turbidity measurement makes an absolute turbidity value not easily retainable for any sensor manufacturer although proper standards and methods are followed. For example, turbidity values greater than 1.0 NTU measured on real-world samples with different sensors, even from the single manufacturer, could differ significantly. For ultra-low turbidity (less than 0.3 NTU) measurement using the same methods (ISO7027 or EPA 180.1), it is likely that the values from different sensor can agree within 0.05 NTU. As such, the user should choose a calibration method and remain with the same calibration method for consistency.

The use of the solid-state calibration is ideal for this situation. For regulatory reporting, especially for ultra-low level of turbidity measurement, it is recommended that the solid-state calibration traceability to the Formazin standards as specified by the regulatory authority be established.

Specifications

Item	T-CAL 736	T-CAL 737	T-CAL 739
P/N	53229	53227	53228
Sensor Calibrated	LT-736 / LT-736B	LT-737 / LT-737B	LT-739 / LT-739B
Quantity Adapters Included	4	3	3
Turbidities Included (NTU)	0.10 ± 0.05 NTU 8.0 ± 0.8 NTU 25.0 ± 0.4 NTU 600 ± 60 NTU	0.10 ± 0.05 NTU 1.5 ± 0.15 NTU 4.0 ± 0.4 NTU	0.10 ± 0.05 NTU 8.0 ± 0.8 NTU 25.0 ± 2.5 NTU